

## INSTRUCTIONS FOR USE

# MaisSafe™

I.V. CANNULA  
with Injection Port & Wings

### Intended Use:

The intended use of this product is access to the human circulatory system for introduction of fluids or medicament and/or Withdrawal of blood samples. It provides a closed path system of drug container & patients.

### Materials Used:

PTFE/FEP/PUR, PP, PE, POM, Silicon Tube, Stainless Steel & BaSO<sub>4</sub>  
(BaSO<sub>4</sub>: used for radio-opaque lining on catheter)

### Indications:

- Infusion of I.V. Solutions.
- Intermittent intravenous drugs administration.

 Non-Pyrogenic

 Latex Free

### Contra Indication:-

- Product should not be used on patients with known hypersensitivity to any of the materials used.
- Administration of high viscous fluids.
- Large blood transfusion.

 Do Not Reuse

 Medical Device

 Do Not Resterilize

### Instructions for use:

- Carefully select and aseptically prepare the site.
- Select suitable size of I.V. Catheter/Cannula & inspect visually to ascertain that package is intact.
- Remove catheter from sterile packing and remove its needle cover.
- Grip the catheter from projection provided on hub.
- Perform venipuncture & check for flashback of blood in flashback chamber.
- Advance the catheter into vein and simultaneously withdraw the needle.
- Withdraw the needle completely while pressing the vein just after the tip of catheter into the vein & discard the needle in an appropriate container.
- Connect to the I.V. Infusion line.
- Cover the puncture site with sterile dressing.
- Perform routine monitoring & venipuncture site maintenance according to medical norms.

 **Manufactured by:**  
**Mais India Medical Devices Pvt. Ltd.**  
525 P, Sector-37, Pace City-II,  
Gurgaon, Haryana-122001 (INDIA)

 **EC REP**  
**Obelis S.A.**  
Boulevard Général Wahis 53,  
1030 Brussels, Belgium.

## INSTRUCTIONS FOR USE

# MaisSafe™

I.V. CANNULA  
with Injection Port & Wings

### Medical condition:

- To read "instruction for use" carefully before use of the product.
- The use of the product is restricted to a qualified doctor or a paramedic staff

### Warnings:

- The product should not be reprocessed.
- Improper transport and handling may cause structural and/or functional damage to device or packaging.
- The product is non-toxic, sterile & non-pyrogenic.
- Do not Clean or resterilize.
- Do not expose to heat or direct sunlight.
- The product should be used immediately after opening the packaging.
- If the device is reused, then it may cause bacterial infection to the patient.
- **MAIS INDIA disclaims any responsibility for possible consequences resulting from improper use.**

### Caution:

- Never try to reinsert a partially or completely withdrawn needle.
- Do not use if package is damaged.
- Discard after single use.

**Target Age Group:** For All age groups

**Duration of Contact of Device :** Up to 72 Hours for best performance.

**Storage condition:** Cool & Dry Place

**Device Life:** 05 Years

**CE 0123**

 CE Notified Body No.

 Catalogue Number

 Lot No./ Batch No.

 Manufacturer

 Date of Manufacturing

 Use By / Expiry Date

 Consult Instructions for Use

 Temperature limit  
(Upper 40°C and lower 10°C)

 Country of manufacture

 Keep Away From Sunlight

 Humidity limitation  
(Upper 60% & Lower 40%)

 Keep Dry

 Caution

 Do Not Use If Package is Damaged

 Sterilized By Ethylene Oxide

 Authorised Representative  
In The European Community

 Single Sterile Barrier System

 Unique Device Identifier

 Product Manual (Booklet)